

UMPQUA Research Company

Analytical Services
ORELAP ID #OR100031

626 NE Division St - PO Box 609
Myrtle Creek, OR 97457
541-863-5201

FILLING & SHIPPING INSTRUCTIONS FOR LEAD and COPPER

EPA 200.8

NOTE: DO NOT ALLOW WATER TO RUN PRIOR TO SAMPLING

SAMPLING INSTRUCTIONS:

- 1) Take sample first thing in the morning or after water has been sitting for a minimum of 6 hours.
- 2) Take sample from the COLD water tap.
- 3) Collect sample immediately upon opening faucet: **DO NOT ALLOW WATER TO RUN PRIOR TO SAMPLING.**

SHIPPING INSTRUCTIONS:

- 4) Cap tightly and pack samples securely for transport to the laboratory.

SAMPLE COLLECTOR: Please fill out required information and return, with sample, to PWS personnel. Do not return this form to the lab.

SAMPLE LOCATION	SAMPLE INFORMATION
CUSTOMER NAME:	DATE SAMPLED:
ADDRESS:	TIME SAMPLED:
SAMPLE ID#: PWS ID#:	WHERE SAMPLE TAKEN (Circle One): Kitchen Faucet Bathroom Faucet OTHER: _____